



CERTIFICATE OF ACCREDITATION

ANSI National Accreditation Board
11617 Coldwater Road, Fort Wayne, IN 46845 USA

This is to certify that

Kanawha Scales and Systems, Inc.
35 Haas Drive
Englewood, OH 45322

has been assessed by ANAB and meets the requirements of international standard

ISO/IEC 17025:2017

while demonstrating technical competence in the field of

CALIBRATION

Refer to the accompanying Scope of Accreditation for information regarding the types of activities to which this accreditation applies

L1166.02-1

Certificate Number


ANAB Approval

Certificate Valid Through: 03/27/2021
Version No. 005 Issued: 02/19/2020



This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

Kanawha Scales and Systems, Inc.

35 Haas Drive
Englewood, OH 45322
Alex Padon
304-755-8321

CALIBRATION

Valid to: March 27, 2021

Certificate Number: L1166.02-1

Mass and Mass Related

Table with 4 columns: Parameter/Equipment, Range, Expanded Uncertainty of Measurement (+/-) 3, Reference Standard, Method, and/or Equipment. Rows include Class I, II, III, and III L scales.

Calibration and Measurement Capability (CMC) is expressed in terms of the measurement parameter, measurement range, expanded uncertainty of measurement and reference standard, method, and/or equipment. The expanded uncertainty of measurement is expressed as the standard uncertainty of the measurement multiplied by a coverage factor of 2 (k=2), corresponding to a confidence level of approximately 95%.

Notes:

- 1. On-site calibration service is available for this parameter, since on-site conditions are typically more variable than those in the laboratory, larger measurement uncertainties are expected on-site than what is reported on the accredited scope.
2. Industrial Scales include but not limited to lab balances, bench scales, floor scales, crane/hanging scales, tank and hopper scales, forklift scales and vehicle scales.
3. The CMCs for balances and scales are highly dependent on the resolution of the unit under test. The CMCs presented here do not include the resolution of the unit under test. The resolution will be included in the reported uncertainty at the time of calibration.
4. This scope is formatted as part of a single document including Certificate of Accreditation No. L1166.02-1

Signature of R. D. ... Vice President

