



# CERTIFICATE OF ACCREDITATION

## The ANSI National Accreditation Board

Hereby attests that

**Kanawha Scales and Systems, Inc.**

**35 Haas Drive  
Englewood, OH 45322**

Fulfills the requirements of

**ISO/IEC 17025:2017**

In the field of

**CALIBRATION**

This certificate is valid only when accompanied by a current scope of accreditation document.  
The current scope of accreditation can be verified at [www.anab.org](http://www.anab.org).

A handwritten signature in black ink, appearing to read 'R. Douglas Leonard Jr.', is positioned above a horizontal line.

R. Douglas Leonard Jr., VP, PILR SBU

Expiry Date: 27 March 2023

Certificate Number: L1166.02-1



This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017.  
This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory  
quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).

**SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017**

**Kanawha Scales and Systems, Inc.**

35 Haas Drive  
Englewood, OH 45322  
Alex Padon  
304-755-8321

**CALIBRATION**

Valid to: **March 27, 2023**

Certificate Number: **L1166.02-1**

**Mass and Mass Related**

<b>Parameter/Equipment</b>	<b>Range</b>	<b>Expanded Uncertainty of Measurement (+/-) <sup>3</sup></b>	<b>Reference Standard, Method, and/or Equipment</b>
Class I, Unmarked and High Precision Lab Balances <sup>1</sup>	(0.002 to 20) g (20.1 to 50 000) g	0.000 61 % of Applied Load 0.000 31 % of Applied Load	ASTM E617 Class 1 Weights and NIST Handbook 44 utilized for the calibration of the Weighing System
Class II, Unmarked and High Precision Balances & Scales <sup>1</sup>	(0.003 to 10) g (10.1 to 500) g	0.000 91 % of Applied Load 0.000 63 % of Applied Load	ASTM E617 Class 2 Weights and NIST Handbook 44 utilized for the calibration of the Weighing System
Class III, Unmarked & Equivalent Industrial Scales <sup>1,2</sup>	(0.001 to 500 000) lb (0.001 to 100 000) kg	0.01 % of Applied Load 0.01 % of Applied Load	NIST Class F and/or ASTM E617 Class 6 Weights and NIST Handbook 44 utilized for the calibration of the Weighing System
Class IIIL Vehicle and Hopper Scales <sup>1</sup>	(500 to 30 000) lb (30 001 to 500 000) lb	0.017 % of Applied Load 0.038 % of Applied Load	NIST Class F and/or ASTM E617 Class 6 Weights and NIST Handbook 44 utilized for the calibration of the Weighing System

Calibration and Measurement Capability (CMC) is expressed in terms of the measurement parameter, measurement range, expanded uncertainty of measurement and reference standard, method, and/or equipment. The expanded uncertainty of measurement is expressed as the standard uncertainty of the measurement multiplied by a coverage factor of 2 ( $k=2$ ), corresponding to a confidence level of approximately 95%.

Notes:

1. On-site calibration service is available for this parameter, since on-site conditions are typically more variable than those in the laboratory, larger measurement uncertainties are expected on-site than what is reported on the accredited scope.
2. Industrial Scales include but not limited to lab balances, bench scales, floor scales, crane/hanging scales, tank and hopper scales, forklift scales and vehicle scales.
3. The CMCs for balances and scales are highly dependent on the resolution of the unit under test. The CMCs presented here do not include the resolution of the unit under test. The resolution will be included in the reported uncertainty at the time of calibration.
4. This scope is formatted as part of a single document including Certificate of Accreditation No. L1166.02-1



R. Douglas Leonard Jr., VP, PILR SBU

